

Ozone absorption coefficients and haze corrections for total ozone measurements with Dobson spectrophotometer*

K. R. RAMANATHAN, P. D. ANGREJI and G. M. SHAH

Physical Research Laboratory, Ahmedabad

(Received 7 May 1965)

ABSTRACT. From laboratory measurements of absorption in an ozone cell with a Dobson spectrophotometer and their comparison with the results of observations made with atmospheric ozone, Dobson has concluded that the absorption coefficients of ozone (from Vigroux' work) for the B and C wavelengths are too high and have to be revised. Observations made at Ahmedabad confirm this. Accepting that the difference of ozone absorption coefficients for A and D is 1.388 and assuming that by using the AD difference method the haze correction is eliminated, we determine the total ozone amount. Combining this with measurements of apparent total ozone with A wave-lengths alone, the haze correction is calculated. With this haze correction and known ozone amount, the absorption coefficients for C and B wave lengths can be calculated. The same absorption coefficients are obtained under varying conditions of haze. They are nearly the same as those obtained by Dobson from his laboratory measurements with an ozone cell.

When the absorption coefficients of ozone as determined by Vigroux are used to calculate the amount of ozone in the atmosphere from measurements with the Dobson spectrophotometer, the ozone values found with the different wave-length pairs A, B, C and D are not identical.

This may be due either to wrong values of the absorption coefficients used or the presence of some other absorbing gas in the atmosphere. In order to test this, Dobson made measurements of ozone absorption in a silica vessel placed immediately above the inlet slit of the standard ozone spectrophotometer, and calculated the relative absorptions of ozone for the different wave-lengths. He also calculated the absorption coefficients for the different wave-length pairs from atmospheric measurements made at Oxford (1957) and at Edmonton (1958) assuming that the difference between the absorption coefficients for A and D wave-lengths was equal to 1.388. The figures in the different rows of Table I give respectively the absorption co-

efficients of A, B, C and D, (1) according to Vigroux as determined from his laboratory measurements, (2) from atmospheric measurements at Oxford (1957) and (3) from atmospheric measurements at Oxford, Edmonton (1958) and - Ahmedabad. Dobson suggested that if the absorption coefficients given in row 2 were used to get the ozone values from observations on different wave-lengths, the results would be identical.

It had been noticed in Ahmedabad that the ozone values determined from AD, AC and CD were not identical if Vigroux's values for the absorption coefficients of C were used. An attempt was made to use the new absorption coefficients given in row (2) of Table 1 to calculate the ozone amounts. There was much better agreement.

The difference of absorption coefficients $(\alpha - \alpha')_A - (\alpha - \alpha')_D$ for AD wave-lengths was assumed to be constant at 1.388 and the ozone amount x_{AD} was calculated using the standard difference formula. Assuming

*Presented at the International Ozone Symposium held at Albuquerque, New Mexico, September 1964

TABLE 1
Differences of absorption coefficients $(\alpha-\alpha)_{\lambda}$ of ozone for different wavelength pairs

| | A | B | C | D | A—D |
|--|-------|-------|-------|-------|-------|
| From Vigroux' observations | 1.762 | 1.214 | 0.865 | 0.374 | 1.388 |
| As determined from Dobson's laboratory measurements with ozone cell at Oxford assuming Vigroux' value of 1.388 for A—D | 1.742 | 1.142 | 0.808 | 0.354 | 1.388 |
| From atmospheric observations at (assuming A—D=1.388) | | | | | |
| Oxford | 1.741 | 1.144 | 0.791 | 0.353 | 1.388 |
| Edmonton | 1.743 | 1.156 | 0.804 | 0.355 | 1.388 |
| Ahmedabad | 1.742 | 1.143 | 0.799 | 0.353 | 1.388 |
| Kodaikanal | 1.744 | — | 0.806 | 0.354 | 1.388 |

α and α' refer to the following pairs of wave-lengths used in ozone measurements

| | λ | λ' | | λ | λ' |
|---|-------------------|-------------------|---|-------------------|-------------------|
| A | 3055 \AA | 3254 \AA | C | 3114 \AA | 3324 \AA |
| B | 3088 \AA | 3291 \AA | D | 3176 \AA | 3398 \AA |

$(\alpha-\alpha')_A$ to be 1.742 which is the mean of the values given in lines (2), (3) and (4) of Table 1, the apparent ozone values x_A were also calculated using the formula

$$p x_A (\alpha-\alpha')_A = \Delta N_A - (\beta-\beta')_A \cdot (p/p_0) \cdot m$$

in which the correction due to haze has been neglected. Writing $(x_A - x_{AD}) (\alpha-\alpha')_A = (\delta-\delta')_A$, $m (\delta-\delta')_A$ would represent the haze correction to be applied to ΔN_A . Other evidence had shown that the haze correction was practically independent of wave-length when the wave-length difference did not exceed 200 \AA . The haze correction thus obtained from x_{AD} and x_A , together with the ozone value x_{AD} , was then used to determine the absorption coefficients of A and D separately. The same method was also used to calculate the absorption coefficients of B and C. Line (5) of Table 1 shows the absorption coefficients of A, B, C and D as determined from the atmospheric measurements at Ahmedabad. The ozone amounts

and absorption coefficients on individual days and the corresponding haze corrections are given in Table 2. Table 3 gives the ozone values at Ahmedabad calculated by AD, AC and CD methods, and also the haze corrections to be applied to x_A , x_C and x_D to make them agree with one another and also with x_{AD} , x_{AC} and x_{CD} .

Similar calculations made from observations made at Kodaikanal where the mean atmospheric pressure is 772 mb on A, D and C wave-lengths gave equally good agreement when the differential absorption coefficient for C was taken to be 0.806 and the haze corrections for A, D and C wave-lengths were assumed to be the same.

The good agreement of the calculated values shows that it is possible not only to get consistent values of ozone amounts using observations with different wave-lengths, but also to determine the day-to-day variations of aerosols from Dobson spectrophotometer observations, if we use the new absorption

TABLE 2
Ozone absorption coefficients and haze corrections
Measurements at Ahmedabad (Direct Sun) — A, D and C

| Date | μ_{AD} | μ_O | x_{AD} | x_A | $(\delta-\delta')_A$ | $(\alpha-\alpha')_A$ | $(\alpha-\alpha')_D$ | $(\alpha-\alpha')_O$ |
|---------|------------|---------|----------|-------|----------------------|----------------------|----------------------|----------------------|
| 3-7-63 | 2.106 | 2.081 | 0.260 | 0.264 | +0.007 | 1.741 | 0.351 | 0.809 |
| 2-11-63 | 2.611 | 2.653 | 0.237 | 0.244 | 0.012 | 1.740 | 0.349 | 0.796 |
| 23-1-64 | 3.173 | 3.144 | 0.245 | 0.248 | 0.005 | 1.741 | 0.351 | 0.799 |
| 29-1-64 | 1.424 | 1.416 | 0.221 | 0.223 | 0.003 | 1.743 | 0.355 | 0.799 |
| 5-2-64 | 2.446 | 2.374 | 0.254 | 0.261 | 0.012 | 1.743 | 0.354 | 0.792 |
| 7-3-64 | 1.925 | 1.968 | 0.236 | 0.240 | 0.006 | 1.744 | 0.353 | 0.798 |
| 9-3-64 | 1.709 | 1.736 | 0.231 | 0.235 | 0.007 | 1.743 | 0.353 | 0.800 |
| 13-3-64 | 2.265 | 2.210 | 0.223 | 0.230 | 0.014 | 1.742 | 0.347 | 0.793 |
| 14-3-64 | 1.929 | 1.964 | 0.221 | 0.225 | 0.009 | 1.743 | 0.351 | 0.801 |
| 16-3-64 | 2.141 | 2.190 | 0.232 | 0.237 | 0.009 | 1.740 | 0.353 | 0.796 |
| 24-3-64 | 2.062 | 2.026 | 0.236 | 0.239 | 0.005 | 1.741 | 0.354 | 0.803 |
| 7-4-64 | 2.051 | 2.019 | 0.239 | 0.241 | 0.003 | 1.744 | 0.356 | 0.806 |
| 8-4-64 | 1.896 | 1.871 | 0.244 | 0.244 | 0.000 | 1.744 | 0.352 | 0.800 |
| 15-4-64 | 2.098 | 2.060 | 0.254 | 0.253 | -0.002 | 1.741 | 0.351 | 0.794 |
| 16-4-64 | 1.888 | 1.915 | 0.257 | 0.257 | 0.000 | 1.739 | 0.352 | 0.797 |
| 18-4-64 | 1.621 | 1.645 | 0.251 | 0.253 | 0.003 | 1.743 | 0.351 | 0.802 |
| 20-4-64 | 1.960 | 1.995 | 0.234 | 0.233 | -0.002 | 1.743 | 0.356 | 0.809 |
| 21-4-64 | 2.053 | 2.012 | 0.246 | 0.247 | 0.002 | 1.742 | 0.353 | 0.803 |
| 21-5-64 | 1.550 | 1.566 | 0.259 | 0.259 | 0.000 | 1.744 | 0.356 | 0.813 |
| 21-5-64 | 2.067 | 2.104 | 0.265 | 0.263 | -0.003 | 1.742 | 0.352 | 0.796 |
| 21-5-64 | 2.786 | 2.835 | 0.261 | 0.261 | 0.000 | 1.742 | 0.352 | 0.792 |
| 25-5-64 | 1.166 | 1.172 | 0.259 | 0.262 | 0.005 | 1.741 | 0.354 | 0.801 |
| 25-5-64 | 2.244 | 2.202 | 0.269 | 0.272 | 0.005 | 1.743 | 0.355 | 0.801 |
| 26-5-64 | 1.279 | 1.288 | 0.262 | 0.264 | 0.003 | 1.741 | 0.352 | 0.793 |
| 26-5-64 | 2.487 | 2.537 | 0.269 | 0.272 | 0.006 | 1.740 | 0.351 | 0.790 |
| 26-5-64 | 2.768 | 2.721 | 0.266 | 0.270 | 0.007 | 1.743 | 0.353 | 0.791 |
| | | | | | Mean | 1.742 | 0.353 | 0.799 |

(Direct Sun) A, D, B and C

| Date | μ_{AD} | μ_B | x_{AD} | x_A | $(\delta-\delta')_A$ | $(\alpha-\alpha')_A$ | $(\alpha-\alpha')_D$ | $(\alpha-\alpha')_O$ | $(\alpha-\alpha')_B$ |
|----------|------------|---------|----------|-------|----------------------|----------------------|----------------------|----------------------|----------------------|
| 3-10-64 | 2.842 | 2.766 | 0.243 | 0.248 | 0.009 | 1.742 | 0.352 | 0.810 | 1.143 |
| 4-10-64 | 2.178 | 2.163 | 0.248 | 0.256 | 0.013 | 1.742 | 0.352 | 0.793 | 1.148 |
| 6-10-64 | 2.259 | 2.333 | 0.254 | 0.257 | 0.005 | 1.742 | 0.352 | 0.798 | 1.148 |
| 7-10-64 | 2.737 | 2.842 | 0.247 | 0.252 | 0.009 | 1.742 | 0.352 | 0.799 | 1.135 |
| 12-10-64 | 2.045 | 2.167 | 0.238 | 0.243 | 0.009 | 1.743 | 0.357 | 0.805 | 1.143 |
| | | | | | Mean | 1.742 | 0.353 | 0.801 | 1.143 |

TABLE 3

Ahmedabad

$$\text{Haze correction } (\delta - \delta')_{\lambda} = (x_{\lambda} - x_{AD}) (\alpha - \alpha')_{\lambda}$$

| Date | Ozone amount (in D. U.) | | | Haze correction | | | Remarks | |
|---------|-------------------------|----------|-------------------|-----------------|------|------|---------|------------------|
| | μ_{AD} | x_{AD} | $x_{A_{\lambda}}$ | x_{CD} | A | C | | D |
| 3-7-63 | 2.106 | 260 | 260 | 262 | 7.0 | 7.2 | 6.3 | — |
| 2-11-63 | 2.611 | 237 | 238 | 238 | 12.2 | 11.2 | 10.9 | — |
| 23-1-64 | 3.173 | 245 | 246 | 245 | 5.2 | 4.8 | 4.4 | — |
| 29-1-64 | 1.424 | 221 | 222 | 220 | 3.5 | 2.8 | 3.2 | — |
| 5-2-64 | 2.446 | 254 | 256 | 251 | 12.2 | 11.2 | 12.1 | Haze |
| 7-3-64 | 1.925 | 236 | 237 | 235 | 7.0 | 5.6 | 5.7 | Haze |
| 9-3-64 | 1.709 | 231 | 232 | 231 | 7.0 | 7.2 | 6.7 | Haze |
| 13-3-64 | 2.265 | 223 | 224 | 221 | 12.2 | 11.6 | 12.2 | Haze |
| 14-3-64 | 1.929 | 221 | 220 | 221 | 7.0 | 8.4 | 8.1 | Haze |
| 16-3-64 | 2.141 | 232 | 233 | 230 | 8.7 | 8.0 | 8.7 | Haze |
| 24-3-64 | 2.062 | 236 | 235 | 237 | 5.2 | 5.6 | 5.0 | Haze |
| 7-4-64 | 2.051 | 239 | 239 | 240 | 3.5 | 4.4 | 3.5 | Thin Ci |
| 8-4-64 | 1.896 | 244 | 244 | 243 | 0.0 | 0.8 | 1.2 | Thin Ci |
| 15-4-64 | 2.098 | 254 | 256 | 252 | 1.7 | -3.6 | -2.8 | Low haze |
| 16-4-64 | 1.888 | 257 | 257 | 256 | 0.0 | -0.8 | -0.5 | Low haze |
| 18-4-64 | 1.621 | 251 | 251 | 253 | 3.5 | 3.6 | 2.1 | Haze |
| 20-4-64 | 1.960 | 234 | 233 | 237 | -1.7 | 0.0 | -1.6 | Low level haze |
| 21-4-64 | 2.053 | 246 | 246 | 248 | 1.7 | 2.8 | 1.8 | Slight haze |
| 21-5-64 | 1.550 | 259 | 257 | 265 | 0.0 | 3.2 | 0.4 | } Low level haze |
| | 2.067 | 265 | 266 | 263 | -3.5 | -4.0 | -3.4 | |
| | 2.786 | 261 | 263 | 257 | 0.0 | -2.4 | -0.7 | |
| 25-5-64 | 1.166 | 259 | 259 | 259 | 5.2 | 5.2 | 5.1 | } Dust haze |
| | 2.244 | 269 | 269 | 269 | 5.2 | 5.6 | 5.3 | |
| 26-5-64 | 1.279 | 262 | 264 | 258 | 3.5 | 1.2 | 2.5 | } Dust haze |
| | 2.487 | 269 | 272 | 264 | 5.2 | 3.2 | 5.0 | |
| | 2.768 | 266 | 269 | 261 | 7.0 | 4.8 | 6.7 | |

coefficients of ozone. It is interesting to note that low level haze (such as occurred on 15-4-64, 20-4-64 and 21-5-64) tends to reduce the haze correction and even to make it negative.

We are thankful to Dr. M. K. Vainu Bappu, Director of the Astrophysical Observatory, Kodaikanal and his staff for sending us the observations made with the Dobson spectrophotometer at Kodaikanal.

REFERENCES

- Dobson, G. M. B. 1963 *Quart. J.R. met. Soc.*, 89, p. 409.
 Vigroux, E. 1953 *Ann. Phys.*, 12, 8, p. 1.